



CERTIFICATE OF ANALYSIS

STP® pHox® Plus M Ionized Magnesium Sensor
Lot Number 24305026

Part Number 45172
Exp Date 2026-01-31
Date of Manufacture 2024-10-31
UDI# (01)10385480451724(11)241031(17)260131(10)24305026

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	9.70 - 14.00	Pass
Std A mV	-100.00 - +60.00	Pass
pHOx + M CtL 1	0.80 - 1.10	Pass
pHOx + M CtL 3	0.25 - 0.37	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

11/18/2024

Date