



CERTIFICATE OF ANALYSIS

Bioprofile Flex K Sensor
Lot Number 24291056

Part Number 42460
Exp Date 2026-01

	<u>Pass/Fail</u>
Slope	Pass
Std A mV	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

Approval

QC Inspector
Title

11/20/2024
Date