



CERTIFICATE OF ANALYSIS

Nova CRT 16™ Creatinine Sensor
Lot Number 24316038

Part Number 17360

Exp Date 2026-11-11

Date of manufacture 2024-11-11

UDI# (01)00385480173605(11)241111(17)261111(10)24316038

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	25.0 - 180.0	Pass
Std A mV	< 500.00	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02453 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector

Title

11/20/2024

Date