



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample
Lot Number 24320078

Ref Number 53463
Exp Date 2026-05-04
Date of Manufacture 2024-11-04
UDI# (01)00385480534635(11)241104(17)260504(10)24320078

Cal A	L/N	24317051	
		<u>Assigned Value*</u>	<u>Result</u>
pH		7.354	N/A
pCO2 mmHg		24.7	N/A

Cal B	L/N	24309061	
		<u>Spec</u>	<u>Pass/Fail</u>
pH		6.837 - 6.843	Pass
pO2 mmHg		167.0	N/A

Flush	L/N	24309073	
		<u>Assigned Value*</u>	<u>Result</u>
pCO2 mmHg		49.4	N/A
pO2 mmHg		101.3	N/A

Reference	L/N	24309075	
		<u>Spec</u>	<u>Pass/Fail</u>
KCl		1.8 - 2.2	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

11/25/2024
Date