



**CERTIFICATE OF ANALYSIS**

Nova 1 Calibrator Pack  
Lot Number 302572

Ref Number 15646  
Exp Date 2025-01-25  
Date of Manufacture 2023-01-25  
UDI# (01)00385480156462(11)230125(17)250125(10)302572

Cal A	L/N	23025072	Cal B	L/N	23025072	Reference	L/N	23025072	
		<u>Spec</u>			<u>Spec</u>			<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.0 - 141.0	Pass	Na mmol/L	59.4 - 60.6	Pass	KCl mol/L	1.8 - 2.2	Pass	
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.8 - 10.2	Pass				

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

*Auishi M*

Approval

QC Inspector  
Title

2/15/2023  
Date