



CERTIFICATE OF ANALYSIS

Nova CRT 16™ Creatinine Membrane Kit
Lot Number 24338014

Part Number 28241
Exp Date 2026-03-03
Date of manufacture 2024-12-03
UDI# (01)00385480282413(11)241203(17)260303(10)24338014

	<u>Spec*</u>	<u>Mean Result</u>
Slope	> 45.0	111.5
Std A mV	< 500.00	384.54

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "Sean Quinn", written over a horizontal line.

Approval

QC Inspector
Title

12/09/2024
Date