



CERTIFICATE OF ANALYSIS

Allegro UACR Test Cartridge
Lot Number 24344017

Ref Number 54214
Exp Date 2025-12-09
Date of Manufacture 2024-12-09
UDI# (01)00385480542142(11)241209(17)251209(10)24344017

	Ctl Level 1	Pass/Fail		Ctl Level 2	Pass/Fail
	Spec			Spec	
Albumin mg/L	29 - 53	Pass	Albumin mg/L	134 - 210	Pass
Creatinine mg/dL	60 - 104	Pass	Creatinine mg/dL	168 - 284	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/12/2024

Date