



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane  
Lot Number 24340013

Part Number 48552

Exp Date 2026-03-05

Date of Manufacture 2024-12-05

UDI# (01)00385480485524(11)241205(17)260305(10)24340013

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	26.73
pHox Ultra/CCX Ctl L4	73 - 87	79
pHox Ultra/CCX Ctl L5	255 - 305	295

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/6/2024

Date