



CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pO2 Sensor
Lot Number 24339018

Ref Number 21521
Exp Date 2026-12-04

Date of Manufacture 2024-12-04

UDI# (01)00385480215213(11)241204(17)261204(10)24339018

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector
Title

12/18/2024
Date