



CERTIFICATE OF ANALYSIS

STP Prime Plus Reference Cartridge
Lot Number 24332054

Ref Number 57823

Exp Date 2026-05-27

Date of Manufacture 2024-11-27

UDI# (01)00385480578233(11)241127(17)260527(10)24332054

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	139.3 - 147.3	Pass
K, mmol/L	3.73 - 4.23	Pass
CL, mmol/L	123.4 - 132.4	Pass
External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na, mmol/L	134.3 - 142.3	Pass
K, mmol/L	3.73 - 4.23	Pass
CL, mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

12/18/2024

Date