

## **CERTIFICATE OF ANALYSIS**

## Nova StatStrip Glucose Control Solution Level 2

Nova Reference 54521 LifeScan Reference 02315502 Lot Number 0424326242 Date of Manufacture 2024-11-21 Exp Date 2026-11-21 UDI# (01)00385480545211(11)241121(17)261121(10)0424326242

Spec Glucose mg/dL 105 - 125

<u>Mean</u> Result 5 116

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Mull

Approval

<u>QC Inspector</u> Title

<u>12/19/2024</u> Date