



## CERTIFICATE OF ANALYSIS

Nova Lactate Plus Control Solution Level 1  
Lot Number 3824330301

Ref Number 40815

Exp Date 2027-05-25

Date of Manufacture 2024-11-25

UDI# (01)00385480408158(11)241125(17)270525(10)3824330301

	<u>Spec</u>	<u>Mean Result</u>
Lactate mmol/L	1.3 - 1.5	1.3

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in purple ink, appearing to read "M. Houghton", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector  
Title

12/20/2024  
Date