



CERTIFICATE OF ANALYSIS

Prime ES plus Cal Cartridge 200 Sample
Lot Number 24352120

Ref Number 62226
Exp Date 2026-05-06
Date of Manufacture 2024-11-06
UDI# (01)00385480622264(11)241106(17)260506(10)24352120

| Cal A L/N 24311037 | | | Cal B L/N 24311038 | | | Reference L/N 24312025 | | |
|--------------------|---------------|------------------|--------------------|--------------|------------------|------------------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| Na mmol/L | 134.1 - 135.9 | Pass | Na mmol/L | 76.5 - 77.5 | Pass | KCl mol/L | 1.8 - 2.2 | Pass |
| K mmol/L | 4.02 - 4.08 | Pass | K mmol/L | 9.80 - 10.20 | Pass | | | |
| Cl mmol/L | 114.1 - 115.9 | Pass | Cl mmol/L | 44.6 - 45.5 | Pass | | | |
| iCa mmol/L | 0.99 - 1.01 | Pass | iCa mmol/L | 1.98 - 2.02 | Pass | | | |
| iMg mmol/L | 0.38 - 0.48 | Pass | iMg mmol/L | 1.48 - 1.52 | Pass | | | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

12/24/2024

Date