



## CERTIFICATE OF ANALYSIS

Stat Profile® pHox Ultra™ / CCX™ SO2 Sensor  
Lot Number 24341070

Part Number 48591  
Exp Date 2027-12-06  
Date of manufacture 2024-12-06  
UDI#(01)00385480485913(11)241206(17)271206(10)24341070

	<u>Spec</u>	<u>Mean Result</u>	<u>Pass/Fail</u>
Slope	7.00 - 14.00	9.22	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. M.", with a red horizontal line drawn across the bottom of the signature.

Approval

QC Inspector  
Title

12/23/2024  
Date