



StatStrip Glucose Control Solution Level 1

Nova Ref Number 54520
LifeScan Ref Number 02315402
Lot Number 0424326241

Date of Manufacture 2024-11-21
Exp Date 2026-11-21
UDI# (01)00385480545204(11)241121(17)261121(10)0424326241

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	57-65	61

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/27/2024
Date