



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane
Lot Number 24352021

Part Number 48552
Exp Date 2026-03-17
Date of Manufacture 2024-12-17
UDI# (01)00385480485524(11)241217(17)260317(10)24352021

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	36.75
pHox Ultra/CCX Ctl L4	73 - 87	80
pHox Ultra/CCX Ctl L5	255 - 305	284

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. [unclear]", written in a cursive style.

Approval

QC Inspector
Title

12/30/2024
Date