



CERTIFICATE OF ANALYSIS

STP Prime Plus Vet Reference Cartridge
Lot Number 24352139

Ref Number 59345
Exp Date 2026-06-17
Date of Manufacture 2024-12-17

Internal QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	139.3 - 147.3	Pass
K mmol/L	3.73 - 4.23	Pass
CL mmol/L	123.4 - 132.4	Pass

External QC Level 4	<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	134.3 - 142.3	Pass
K mmol/L	3.73 - 4.23	Pass
CL mmol/L	118.9 - 127.9	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

12/27/2024
Date