



## CERTIFICATE OF ANALYSIS

STP pHox/pHOx Basic/BiopHOx Air Detector  
Lot Number 24338122

Ref Number 21513  
Date of Manufacture 2024-12-03

	<u>Spec</u>	<u>Pass/Fail</u>
AD1 mV	1400 - 2600	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in blue ink, appearing to read "C. Brown", with a stylized flourish at the end.

Approval

QC Inspector  
Title

12/31/2024  
Date