



## CERTIFICATE OF ANALYSIS

STP pHox /Basic/Plus/L/C/M pH Sensor  
Lot Number 24362028

Ref Number 21522  
Exp Date 2026-12-27  
Date of Manufacture 2024-12-27  
UDI# (01)00385480215220(11)241227(17)261227(10)24362028

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	10.0 - 10.8	Pass
Std A mV	0.00 +/- 110.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

A handwritten signature in blue ink, appearing to read "C. R. ...", with a stylized flourish at the end.

Approval

QC Inspector  
Title

1/7/2025  
Date