



CERTIFICATE OF ANALYSIS

STP Prime CCS Comp Cal Cartridge 300 Sample
Lot Number 24358101

Ref Number 52427
Exp Date 2026-05-25
Date of Manufacture 2024-11-25
UDI# (01)00385480524278(11)241125(17)260525(10)24358101

Cal A L/N 24346019			Cal B L/N 24346020		
	<u>Assigned Value*</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
pH	7.352	N/A	pH	6.837 - 6.843	Pass
pCO2 mmHg	24.9	N/A	pO2 mmHg	167.0	N/A
	<u>Spec</u>	<u>Pass/Fail</u>		<u>Spec</u>	<u>Pass/Fail</u>
Na mmol/L	132.1 - 133.9	Pass	Na mmol/L	68.0 - 69.0	Pass
K mmol/L	3.97 - 4.03	Pass	K mmol/L	9.92 - 10.08	Pass
Cl mmol/L	105.2 - 106.8	Pass	Cl mmol/L	45.6 - 46.4	Pass
iCa mmol/L	1.08 - 1.12	Pass	iCa mmol/L	2.17 - 2.23	Pass
Glu mg/dL	79 - 81	Pass	Glu mg/dL	197 - 203	Pass
Lac mmol/L	1.8 - 2.2	Pass	Lac mmol/L	9.5 - 10.5	Pass

Flush L/N 24330010			Reference L/N 24346018		
	<u>Assigned Value*</u>	<u>Result</u>		<u>Spec</u>	<u>Pass/Fail</u>
pCO2 mmHg	51.3	N/A	KCl	1.8 - 2.2	Pass
pO2 mmHg	100.8	N/A			

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Haneen Aldarwish

Approval

QC Inspector
Title

01/07/2025
Date