



CERTIFICATE OF ANALYSIS

Allegro HbA1c Test Cartridge
Lot Number 25002015

Ref Number 54215
Exp Date 2026-01-02
Date of Manufacture 2025-01-02
UDI# (01)00385480542159(11)250102(17)260102(10)25002015

| Ctl Level 1 | | | Ctl Level 2 | | |
|-------------|-------------|------------------|-------------|-------------|------------------|
| | <u>Spec</u> | <u>Pass/Fail</u> | | <u>Spec</u> | <u>Pass/Fail</u> |
| HbA1c % | 4.9 - 6.7 | Pass | HbA1c % | 8.1 - 11.7 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/6/2025

Date