



CERTIFICATE OF ANALYSIS

Bioprofile 400/Basic/100 Plus Glutamine Membrane
Lot Number 25003041

Part Number 24456
Exp Date 2026-01

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.0 - 60.0	46.0
Biopro Ctl L4	0.82 - 1.32	1.11
Biopro Ctl L5	4.35 - 5.85	5.13

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector
Title

1/10/2025
Date