



CERTIFICATE OF ANALYSIS

Nova Primary Linearity Level 1

Ref Number 63165

Lot Number 24341056

Exp Date 2026-06-06

Date of Manufacture 2024-12-06

UDI# (01)10385480631652(11)241206(17)260606(10)24341056

Level 1

	<u>Spec</u>	<u>Pass/Fail</u>
Glu mg/dL	481 - 531	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

1/10/2025

Date