



## CERTIFICATE OF ANALYSIS

STP pHox Plus K Sensor  
Lot Number 24355013

Part Number 27811

Exp Date 2026-03-20

Date of Manufacture 2024-12-20

UDI# (01)00385480278119(11)241220(17)260320(10)24355013

	<u>Spec*</u>	<u>Pass/Fail</u>
Slope	$\geq 9.60$	Pass
Std A mV	$0 \pm 60.00$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Sandang", written in a cursive style.

Approval

QC Inspector  
Title

01/10/2025  
Date