



CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Calcium Sensor
Lot Number 24347027

Part Number 48581

Exp Date 2026-03-12

Date of Manufacture 2024-12-12

UDI# (01)00385480485814(11)241212(17)260312(10)24347027

| | <u>Spec</u> | <u>Pass/Fail</u> |
|-------------|--------------|------------------|
| Slope | 8.50 - 12.50 | Pass |
| Std C mV | 0.00 ± 50.00 | Pass |
| pHox Ctl L4 | 1.03 - 1.19 | Pass |
| pHox Ctl L5 | 1.51 - 1.75 | Pass |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/10/2025

Date