



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane
Lot Number 25007005

Part Number 48554
Exp Date 2026-04-07
Date of Manufacture 2025-01-07
UDI# (01)00385480485548(11)250107(17)260407(10)25007005

	<u>Spec*</u>	<u>Mean Result</u>
Slope	≥ 15.00	37.59
pHox Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.2
Lin Soln Set A Level 1	14.0 - 20.0	18.5

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. M. ...", written over a horizontal line.

Approval

QC Inspector
Title

1/16/2025
Date