



CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane
Lot Number 25007003

Part Number 48552
Exp Date 2026-04-07
Date of Manufacture 2025-01-07
UDI# (01)00385480485524(11)250107(17)260407(10)25007003

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	34.91
pHox Ultra/CCX Ctl L4	73 - 87	79
pHox Ultra/CCX Ctl L5	255 - 305	289

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "John", written in a cursive style.

Approval

QC Inspector
Title

1/16/2025
Date