



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Glucose Membrane  
Lot Number 25007004

Part Number 48552  
Exp Date 2026-04-07  
Date of Manufacture 2025-01-07  
UDI# (01)00385480485524(11)250107(17)260407(10)25007004

	<u>Spec*</u>	<u>Mean Result</u>
Slope	11.00 - 45.00	36.09
pHox Ultra/CCX Ctl L4	73 - 87	80
pHox Ultra/CCX Ctl L5	255 - 305	293

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "David A. Rosen".

Approval

QC Inspector  
Title

1/21/2025  
Date