



## CERTIFICATE OF ANALYSIS

BioProfile Cell Density Calibrator  
Lot Number 24349002

Ref Number 43034  
Exp Date 2026-03

	<u>Spec</u>	<u>Pass/Fail</u>
Particles/mL	$23.0 \times 10^5 - 27.0 \times 10^5$	Pass

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to read "M. Hughes", with a long, sweeping horizontal line extending to the right.

Approval

QC Inspector  
Title

1/22/2025  
Date