



CERTIFICATE OF ANALYSIS

Allegro UACR Test Cartridge
Lot Number 25020001

Ref Number 54214
Exp Date 2026-01-20
Date of Manufacture 2025-01-20
UDI# (01)00385480542142(11)250120(17)260120(10)25020001

	Ctl Level 1	Pass/Fail		Ctl Level 2	Pass/Fail
	Spec			Spec	
Albumin mg/L	29 - 53	Pass	Albumin mg/L	134 - 210	Pass
Creatinine mg/dL	60 - 104	Pass	Creatinine mg/dL	168 - 284	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/27/2025

Date