



## CERTIFICATE OF ANALYSIS

STP pHox Ultra/CCX Lactate Membrane  
Lot Number 25022024

Part Number 48554  
Exp Date 2026-04-22  
Date of Manufacture 2025-01-22  
UDI# (01)00385480485548(11)250122(17)260422(10)25022024

	<u>Spec*</u>	<u>Mean Result</u>
Slope	$\geq 15.00$	37.20
pHox Ultra/CCX Ctl L4	1.7 - 2.3	2.0
pHox Ultra/CCX Ctl L5	6.3 - 7.7	7.1
Lin Soln Set A Level 1	14.0 - 20.0	19.2

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

1/27/2025  
Date