



CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ Na Sensor
Lot Number 25010068

Ref Number 48579
Exp Date 2026-07-10

Date of Manufacture 2025-01-10

UDI# (01)00385480485791(11)250110(17)260710(10)25010068

| | <u>Spec</u> | <u>Pass/Fail</u> |
|----------|-----------------|------------------|
| Slope | 9.30 - 11.00 | Pass |
| Std C mV | -10.00 - +70.00 | Pass |
| L4 CtL | 136.0 - 144.0 | Pass |
| L5 CtL | 110.5 - 118.5 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/27/2025

Date