



CERTIFICATE OF ANALYSIS

STP Prime Plus Reference Cartridge
Lot Number 25017081

Ref Number 57823

Exp Date 2026-07-17

Date of Manufacture 2025-01-17

UDI# (01)00385480578233(11)250117(17)260717(10)25017081

Table with 3 columns: Test Level, Spec, Pass/Fail. Rows include Internal QC Level 4 (Na, K, CL) and External QC Level 4 (Na, K, CL).

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U. S. A. in accordance with 21 CFR-Part 820, Quality System Regulation (QSR), Current Good Manufacturing Practice (cGMP) and EN ISO 13485:2016, Medical Devices, Quality Management Systems for in Vitro Diagnostic devices, and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

1/28/2025

Date