



**CERTIFICATE OF ANALYSIS**  
**Nova StatStrip Glucose and  $\beta$ -Ketone Control Solution Level 2**

Nova Reference 46948  
LifeScan Reference 02312901  
Lot Number 0425007302

Exp Date 2027-07-07  
Date of Manufacture 2025-01-07  
UDI# (01)00385480469487(11)250107(17)270707(10)0425007302

	<u>Spec</u>	<u>Mean Result</u>
Glucose mg/dL	105 - 125	112
Ketone mmol/L	0.7 - 1.1	1.1

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: The process for assigning values to Nova Biomedical Quality Control and Linearity Solutions utilizes multiple sources of test data inputs to establish the ranges for each analyte in each lot, which may result in small lot-to-lot variation in insert range values.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector  
Title

01/25/2025  
Date