



## CERTIFICATE OF ANALYSIS

STP® pHox Ultra™ / CCX™ Creatinine Sensor  
Lot Number 25013029

Part Number 48595  
Exp Date 2027-01-13  
Date of Manufacture 2025-01-13  
UDI# (01)00385480485951(11)250113(17)270113(10)25013029

|             | <u>Specs</u> | <u>Pass/Fail</u> |
|-------------|--------------|------------------|
| Slope       | ≥ 5.00       | Pass             |
| pHox Ctl L4 | 0.7 - 1.3    | Pass             |
| pHox Ctl L5 | 5.2 - 7.2    | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

01/29/2025

Date