



CERTIFICATE OF ANALYSIS

Nova Cleaning Solution
Lot Number 25022055

Ref Number 11272
Exp Date 2026-01-22
Date of Manufacture 2025-01-22
UDI# (01)00385480112727(11)250122(17)260122(10)25022055

	<u>Spec</u>	<u>Pass/Fail</u>
NaOCl	1.21 - 1.51%	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi M

Approval

QC Inspector
Title

2/3/2025
Date