



## CERTIFICATE OF ANALYSIS

STAT PROFILE® pHox Ultra™/CCX™ pO2 Sensor  
Lot Number 25022054

Ref Number 48598

Exp Date 2027-01-22

Date of Manufacture 2025-01-22

UDI# (01)00385480485982(11)250122(17)270122(10)25022054

	<u>Spec</u>	<u>Pass/Fail</u>
Slope	-3.00 to -6.00	Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. [unclear]", written in a cursive style.

Approval

QC Inspector  
Title

2/6/2025  
Date