



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 100 Sample
Lot Number 25028040

Ref Number 53104
Exp Date 2026-07-13
Date of Manufacture 2025-01-13
UDI# (01)00385480531047(11)250113(17)260713(10)25028040

| | | | |
|-------|-----------|------------------------|---------------|
| Cal A | L/N | 25022044 | |
| | | <u>Assigned Value*</u> | <u>Result</u> |
| | pH | 7.345 | N/A |
| | pCO2 mmHg | 25.5 | N/A |

| | | | |
|-------|----------|---------------|------------------|
| Cal B | L/N | 25022045 | |
| | | <u>Spec</u> | <u>Pass/Fail</u> |
| | pH | 6.837 - 6.843 | Pass |
| | pO2 mmHg | 167.0 | N/A |

| | | | |
|-------|-----------|------------------------|---------------|
| Flush | L/N | 25013041 | |
| | | <u>Assigned Value*</u> | <u>Result</u> |
| | pCO2 mmHg | 51.7 | N/A |
| | pO2 mmHg | 100.5 | N/A |

| | | | |
|-----------|-----|-------------|------------------|
| Reference | L/N | 25022073 | |
| | | <u>Spec</u> | <u>Pass/Fail</u> |
| | KCl | 1.8 - 2.2 | Pass |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Awishi M

Approval

QC Inspector

Title

02/03/2025

Date