



CERTIFICATE OF ANALYSIS

Nova 8 Ionized Magnesium Sensor
Lot Number 25016024

Ref Number 15265
Exp Date 2026-04-16

Date of Manufacture 2025-01-16

UDI# (01)10385480152652(11)250116(17)260416(10)25016024

	<u>Spec*</u>	<u>Mean</u>
Slope	10.5 - 13.5	12.5
Chem Control L1 mmol/L	0.46 - 0.66	0.49
Set A Level 2 mmol/L	0.78 - 1.02	0.87

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

02/11/2025

Date