



## CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamine Membrane  
Lot Number 25038001

Part Number 42486  
Exp Date 2025-08

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	24.52
Biopro Ctl L4	0.74 - 1.24	0.99
Biopro Ctl L5	4.37 - 5.87	4.92

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in blue ink, appearing to read "David A. Rasmussen".

Approval

QC Inspector

Title

2/11/2025

Date