



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 2

Ref Number 41742

Lot Number 0425017302

Exp Date 2027-07-17

Date of Manufacture 2025-01-17

UDI# (01)00385480417426(11)250117(17)270717(10)0425017302

	Spec	Mean Result
Glucose mg/dL	105-125	115

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Sandang", written over a light gray rectangular background.

Approval

QC Inspector

Title

2/12/2025

Date