



CERTIFICATE OF ANALYSIS

Bioprofile Flex Glutamate Membrane Kit
Lot Number 25038002

Part Number 42484
Exp Date 2025-08

	<u>Spec*</u>	<u>Mean Result</u>
Slope	10.00 - 60.00	44.87
Biopro Ctl L4	0.28 - 0.78	0.50
Biopro Ctl L5	1.78 - 2.38	2.15

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in dark ink, appearing to be "J. Smith", written over a faint horizontal line.

Approval

QC Inspector
Title

2/13/2025
Date