



CERTIFICATE OF ANALYSIS

Nova StatStrip Lactate Control Solution Level 2
Lot Number 2725009302

Ref Number 47554

Exp Date 2027-07-09

Date of Manufacture 2025-01-09

UDI# (01)00385480475549(11)250109(17)270709(10)2725009302

	Spec	Mean Result
Lactate mmol/L	5.9-6.7	5.9

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to be "J. Smith", written over a horizontal line.

Approval

QC Inspector

Title

2/13/2025

Date