OI-922

Page 41 of 67

Rev. BM

Vialed and Packaged Test Strip Release

Test Strips



## **CERTIFICATE OF COMPLIANCE**

This product has been manufactured in compliance with ISO 13485:2016 and Medical Device Single Audit Program (MDSAP) standards and meets Nova Biomedical product specifications.

| Customer:   |                | LIFESCAN  |
|---|----------------|---|
| Description:  |                | TEST STRIPS, STATSTRIP <sup>™</sup> GLUCOSE, PACKAGED 18 BOXES OF 100 COUNT |
| Lifescan Purchase Order No.:  |                | 489030  |
| Catalog Number:   |                | 42214   |
| Nova Biomedical P/N   |                | 54311   |
| One Touch P/N   | *, /           | 02307501  |
| Target Range (mg/dL) SD(mg/dL) /CV (%)  |                |   |
| 30-50   | $\leq$ 4 mg/dL | PASS / FAIL   |
| 51-110  | ≤ 6.0%         | PASS / FAIL   |
| 111-150   | ≤ 5.0%         | PASS/ FAIL  |
| 151-250   | ≤ 5.0%         | PASS FAIL   |
| 251-400   | ≤ 5.0%         | PASS / FAIL   |
| Accuracy Criteria 1: Total within +/-12 mg/dL or +/-12% of reference method $\geq 95\%$ |                | PASS) FAIL  |
| Accuracy Criteria 2: Total within +/-15 mg/dL or +/-15% of reference method $\geq$ 98%  |                | PASS / FAIL   |
| Qty:  |                | 12.096  |
| Lot No.   |                | 0325 041309   |
| Date of Mfg. (DOM):   |                | 2025 / 02 / 10 (YYYY/MM/DD)   |
| Expiration Date:  |                | 2027 / 08 (YYYY/MM)   |
| Unique Device Identifier (UDI):   |                | $ \begin{array}{c ccccccccccccccccccccccccccccccccccc$                      |

Nova Biomedical Corporation 200 Prospect St

Waltham, MA 02454