



**CERTIFICATE OF ANALYSIS**

BioProfile FLEX 2 Chemistry Reference Sensor  
Lot Number 25044066

Part Number 57466  
Exp Date 2026-08-13  
Date of Manufacture 2025-02-13

|                | <u>Spec*</u>  | <u>Pass/Fail</u> |
|----------------|---------------|------------------|
| Na Slope       | 8.70 - 11.80  | Pass             |
| BioPro Ctl L 1 | 70.6 - 86.2   | Pass             |
| BioPro Ctl L 3 | 164.7 - 201.3 | Pass             |

This product is certified to have been manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to test specifications.

\* Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Approval

QC Inspector  
Title

2/19/2025  
Date