



CERTIFICATE OF ANALYSIS

STP Prime+ Cal-Cartridge 100 Sample w. Creat/BUN
Lot Number 25043035

Ref Number 57831

Exp Date 2026-07-30

Date of Manufacture 2025-01-30

UDI# (01)00385480578318(11)250130(17)260730(10)25043035

Table with 3 columns: Cal A L/N 25035023, Cal B L/N 25035024, Cal C L/N 25035025. Rows include HCO3 mmol/L, Na mmol/L, K mmol/L, Cl mmol/L, Ca mmol/L, Mg mmol/L, Glu mg/dL, Lac mmol/L, BUN mg/dL, Creat mg/dL, and tHb g/dL.

Table with 3 columns: Cal D L/N 25035026, Flush L/N 25030006, Reference L/N 25034083. Rows include pH @ 37C, PO2, and KCl.

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector

Title

2/20/2025

Date