



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 300 Sample
Lot Number 25036027

Ref Number 53359
Exp Date 2026-07-14
Date of Manufacture 2025-01-14
UDI# (01)00385480533591(11)250114(17)260714(10)25036027

Cal A	L/N	25014020		
		<u>Assigned Value*</u>	<u>Result</u>	
	pH	7.348	N/A	
	pCO2 mmHg	25.2	N/A	

Cal B	L/N	25014021		
		<u>Spec</u>	<u>Pass/Fail</u>	
	pH	6.837 - 6.843	Pass	
	pO2 mmHg	167.0	N/A	

Flush	L/N	25016029		
		<u>Assigned Value*</u>	<u>Result</u>	
	pCO2 mmHg	50.0	N/A	
	pO2 mmHg	102.3	N/A	

Reference	L/N	25021019		
		<u>Spec</u>	<u>Pass/Fail</u>	
	KCl	1.8 - 2.2	Pass	

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Auishi M

Approval

QC Inspector

Title

02/19/2025

Date