



CERTIFICATE OF ANALYSIS

STP Prime ABG Cal Cartridge 400 Sample
Lot Number 25043068

Ref Number 53463
Exp Date 2026-07-16
Date of Manufacture 2025-01-16
UDI# (01)00385480534635(11)250116(17)260716(10)25043068

| | | | | |
|-------|-----------|------------------------|---------------|--|
| Cal A | L/N | 25034006 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| | pH | 7.356 | N/A | |
| | pCO2 mmHg | 24.6 | N/A | |

| | | | | |
|-------|----------|---------------|------------------|--|
| Cal B | L/N | 25034007 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| | pH | 6.837 - 6.843 | Pass | |
| | pO2 mmHg | 167.0 | N/A | |

| | | | | |
|-------|-----------|------------------------|---------------|--|
| Flush | L/N | 25016029 | | |
| | | <u>Assigned Value*</u> | <u>Result</u> | |
| | pCO2 mmHg | 50.0 | N/A | |
| | pO2 mmHg | 102.3 | N/A | |

| | | | | |
|-----------|-----|-------------|------------------|--|
| Reference | L/N | 25034033 | | |
| | | <u>Spec</u> | <u>Pass/Fail</u> | |
| | KCl | 1.8 - 2.2 | Pass | |

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

Approval

QC Inspector
Title

02/22/2025
Date