

CERTIFICATE OF ANALYSIS

Preheater Deproteinizing Sol'n Lot Number 25030027

Ref Number 12704 7 Exp Date 2026-01-30 Date of Manufacture 2025-01-30 UDI# (01)00385480127042(11)250130(17)260130(10)25030027

NaOCI Spec Pass/Fail Pass

NaOCI 0.70 - 0.90% Pass

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications. All listed analytes are traceable to NIST SRM Materials.

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer.

Auishi.M

Approval

QC Inspector Title

2/25/2025 Date