



CERTIFICATE OF ANALYSIS

Nova StatStrip Glucose Control Solution Level 1

Nova Ref Number 41741

Lot Number 0425028301

Exp Date 2027-07-28

Date of Manufacture 2025-01-28

UDI # (01)00385480417419(11)250128(17)270728(10)0425028301

	Spec	Mean Result
Glucose mg/dL	57 - 65	62

This certifies that this product was manufactured and tested at Nova Biomedical Corporation, Waltham MA 02454 U.S.A. in accordance with ISO 13485:2016 Medical Devices Quality Management Systems Requirements, Medical Device Single Audit Program (MDSAP), and conforms to the indicated test specifications

Note: Acceptance specifications for this part number are lot dependent and subject to change by the manufacturer

A handwritten signature in black ink, appearing to read "Lumina Ahtu".

Approval

QC Inspector
Title

2/24/2025
Date